

# Ohaus Corporation

## Service Notes

**Model**.....FD Series Food Scale  
**Description** ....SW Update Version sr 1.13  
**Release Date** ..... April 15, 2005  
**Issued By**..... Al Go

---

Updated FD Series software to decrease the minimum Tare/Auto-Tare threshold.

Current: Anything 5d and below (ex. FD6: 6kg x 1g scale, it is 5g with LFT off, 10g with LFT on) is considered too close to zero and won't be taken as a tare weight. In food applications, containers that need to be tared may be too light, so this threshold is not acceptable.

This threshold is likewise used for the checkweigh LED and beeper indication. As a battery-saving feature, the LED and beeper will not activate when there is "no weight" (less than 5d) on the pan.

Modification: The 5d threshold is revised to 1d fixed.

Implementation of updated software will be as follows:

1. Software version is now "Sr 1.13"
2. Starting serial numbers:
  - a. FD3 with serial number 0006102-6DG (D = April, G = 2005)
  - b. FD6 with serial number 0006155-6DG
  - c. FD15 with serial number 0006311-6DG
3. Model name or configuration is not changed.
4. Current manuals are not affected.